



Quorum Technologies



Units 1/2 Eden Business Centre, South Stour Avenue, Ashford, Kent TN23 7RS, United Kingdom

## DECLARATION OF CONFORMITY

We Quorum Technologies Limited declare under our sale responsibility that the product below to which this declaration relates is in conformity with the following standards and other normative document(s):

**MODEL:** K850WM **MARKED:** 

**PRODUCT:** Larger Chamber (MEMS) Critical Point Dryer

**DIRECTIVES:** Electromagnetic Compatibility  
E.M.C. 89/336/EEC (U.K. Regulations SI 1992/2372)  
Low Voltage Directive  
L.V.D. 73/23/EEC (U.K. Regulations SI 1994/3260)

### HARMONISED

**STANDARDS:** E.M.C. EN 55011, (Group 1 Class A) Specification for:

Limits & methods of measurement of radio disturbance characteristics of industrial, scientific & medical (ISM) radio-frequency equipment.

EN 50082-1 Specification for:

Electromagnetic compatibility – Generic immunity standard. Part 1. Residential, commercial & light industry.


L.V.D. EN 61010-1 (Class 1) Specification for:

Safety requirements for electrical equipment for measurement, control, and laboratory use. Part 1. General requirements.

**DECLARATION:** This Equipment of this Design and Manufacture conforms with the requirements of the European Directives and European Harmonised Standards.

**MANUFACTURER:** Quorum Technologies Ltd

**ADDRESS:** Units 1/2 Eden Business Centre, South Stour Avenue, Ashford, Kent, TN23 7RS, England.

Approved By:   
P.R. Missing, Technical Director

Date: 8<sup>th</sup> April 2009

Place: Ashford

*This product has been manufactured under a quality system registered to ISO9001*