



CERTIFICATE OF CONFORMITY

MODEL: PP3010T
PRODUCT: CRYO PREPARATION SYSTEM FOR SEM AND FIB/SEM

MANUFACTURER: QUORUM TECHNOLOGIES LTD
ADDRESS: 2 Acorn House, The Broyle, Ringmer,
East Sussex BN8 5NN UK

Conforms to the protection requirements of Council Directive 2004/108/EC, relating to Electromagnetic Compatibility and Council Directive 2006/95/EC, relating to the Low Voltage Directive, by the application of:

EN61326-1: 2006

Electrical equipment for measurement, control and laboratory use.

Part 1: General requirements:

Emissions tested to Class B – Domestic level.

Immunity tested to Class A – Industrial Level.

EN61010-1:2010: Safety requirements for Electrical Equipment for measurement, control & laboratory use. Part 1: General requirements.

Signed:

Date: 10th October 2012

Name: Michael Wombwell

Position: Director